

PCE Americas Inc.
711 Commerce Way
Suite 8
Jupiter
FL-33458
USA
From outside US: +1
Tel: (561) 320-9162
Fax: (561) 320-9176
info@pce-americas.com

PCE Instruments UK Ltd.
Units 12/13
Southpoint Business Park
Ensign way
Hampshire / Southampton
United Kingdom, SO31 4RF
From outside UK: +44
Tel: (0) 2380 98703 0
Fax: (0) 2380 98703 9
info@pce-instruments.com

www.pce-instruments.com/english
www.pce-instruments.com

Manual Analytical Balance PCE-BS 6000



Version 1.0
Date of creation: 21.10.2015
Date of last change: 29.02.2016

Contents

- 1 Introduction 3**
- 2 Safety notes 3**
- 3 Specification 4**
 - 3.1 Technical specifications 4
 - 3.2 Contents of delivery 4
- 4 System description 5**
- 5 Weighing 6**
 - 5.1 General notes 6
 - 5.2 Weighing 6
 - 5.3 Hold function 6
 - 5.4 Tare function 6
- 6 Calibration 7**
- 7 Contact 8**
 - 7.1 PCE Instruments UK 8
 - 7.2 PCE Americas 8

1 Introduction

Thank you for purchasing a tabletop scale from PCE Instruments.

The PCE-BS 6000 tabletop scale offers a compact design, high accuracy and an easy handling at a low price. The weighing platform of the tabletop scale is made of stainless steel, which makes it easy to clean. The scale can be operated by using the integrated mains adaptor or by using 6 x 1.5 V AA batteries. In battery operation, the PCE-BS 6000 is well-suited for mobile application. With its adjustable feet and the integrated water-level, the scale can be arranged properly in no time.

2 Safety notes

Please read this manual carefully and completely before you use the device for the first time. The device may only be used by qualified personnel and repaired by PCE Instruments personnel. There is no warranty of damage or injuries caused by non-observance of the manual.

- The device may only be used in the approved temperature range.
- The case should only be opened by qualified personnel of PCE Instruments.
- The instrument should never be placed with the user interface facing an object (e.g. keyboard side on a table).
- You should not make any technical changes to the device.
- The appliance should only be cleaned with a damp cloth / use only pH-neutral cleaner.
- Do not expose the scale to direct sunlight. This can lead to malfunction.
- Only use the included mains adaptor (9V, 100 mA). In battery operation, only use 1.5 V AA batteries. Replace all batteries at the same time.
- If you do not use the scale for a longer period of time, remove the batteries to prevent damage caused by battery leakage.
- Do not expose the scale to water. Do not use it in areas with high air humidity.
- Keep the scale clean and prevent dust and other particles from reaching the electronics. For cleaning, use a damp cloth. Do not use solvent-containing materials for cleaning.
- Avoid electrostatic discharge when using the scale.
- Avoid vibrations and shocks while the scale is in use.
- Make sure that the scale is always arranged properly. Use the integrated water-level to check.

This manual is published by PCE Instruments without any guarantee.

We expressly point to our general guarantee terms which can be found in our general terms of business.

If you have any questions please contact PCE Instruments.

3 Specification


3.1 Technical specifications

Weighing range	6,000 g
Resolution	1 g
Accuracy	± 3 g
Platform dimensions	160 x 135 mm
Weighing units	Kg, g, lb, oz
Display	LCD
Character height	15 mm
Operating temperature	0 ... +40 °C
Power supply	Mains operation: 9 V / 100 mA Battery operation: 6 x 1.5 V AA batteries
Dimensions	165 x 245 x 70 mm
Weight	Approx. 900 g

3.2 Contents of delivery

- 1 x tabletop scale PCE-BS 6000
- 1 x stainless steel weighing platform
- 1 x mains cable
- 1 x instruction manual

4 System description



5 Weighing

5.1 General notes

Always keep the automatic power-off function in mind.

After 3 minutes of idling time (when there is no weight on the platform), the scale automatically shuts down to save battery power.

Error messages



EEEE Overload (weighing cell load is too high; weighing cell is probably damaged)

Battery voltage too low

5.2 Weighing

1. Turn on the scale by pressing the ON/OFF button.
2. Choose the weighing unit by using the UNIT button. You can select between kg, g, lb and oz.
3. Put the load on the weighing platform and wait for the reading on the display to stabilize.

5.3 Hold function

You can freeze the reading on the display for about 120 s (after the load has been removed).

1. Put the load on the platform.
2. Press the Hold button. A hold indication appears on the display.
3. Remove the load from the platform. The reading now remains on the display of the scale.

If you manually want to unfreeze the reading, press the Zero button. To exit the Hold function, press the Hold button. You get back to normal measuring mode.

5.4 Tare function

You can tare weights by pressing the Zero button (up to half of the weighing range).

6 Calibration

A calibration can be reasonable when the scale is moved a lot or when the environmental conditions vary heavily.

To perform a calibration, follow these steps:

1. Turn on the scale by using the ON/OFF button.
2. Press and hold the CAL button for 2 seconds. The display first shows "CAL" and then a flashing "0".
3. Press the CAL button again. The display now shows the max. load of the scale (6,000 g) in flashing figures.
4. Put a 6 kg calibration weight on top of the platform and press the CAL button one more time. The display now shows "PASS" and you are redirected to normal measuring mode. The calibration has been completed.
5. Turn off the scale and then turn it back on by using the ON/OFF button. Take a few measurements. If you notice or suspect deviations between the measurements, you can repeat the calibration.

7 Contact

If you have any questions about our range of products or measuring instruments please contact PCE Instruments.

7.1 PCE Instruments UK

By post:

PCE Instruments UK Ltd.
Units 12/13 Southpoint Business Park
Ensign Way, Southampton
Hampshire

United Kingdom, SO31 4RF

By phone:

02380 987 035

7.2 PCE Americas

By post:

PCE Americas Inc.
711 Commerce Way
Suite 8
Jupiter
33458 FL
USA

By phone:

561 320 9162

